

Certificate of Performance

TON 3

Product Code KR105

Date 05.03.2020


Expiry Date 05.03.2021

Reagent	Lot Number
TON 1	4313
TON 2	4340
TON 3	4341
Test Method	WC SOP 23
Test Instrument	Konelab 20
Serial Number	P4919028

Std. Value	Lot #	Abs.
0.00		0.003
2.50	TON 100 01 20	0.183
5.00	TON 100 01 20	0.352
7.69	TON 100 01 20	0.513
10.00	TON 100 01 20	0.637


*serial dilution

Calibration Curve



Test	Result	Specification	Pass / Fail
Intercept	0.003	-0.005 to 0.01	Pass
Sensitivity	Abs/ppm		
0 - 2.50 mg/l	0.0720	0.0550 to 0.0800	Pass
2.50 - 5.00 mg/l	0.0676	0.0600 to 0.0800	Pass
5.00 - 7.69 mg/l	0.0599	0.0550 to 0.0800	Pass
7.69 - 10.0 mg/l	0.0537	0.0500 to 0.0700	Pass

Notes
 Store between 2 & 8 °C
 *Stable on Board TON 1 - 1 day
 * TON 2 & TON 3 - 5 days
 Keep Tightly Sealed when not in use.
 Do not return unused portions to container
 Barcode ID- TON 1 103; TON 2 104; TON 3 105
 * Based on results obtained in our Research Facility


 A. Coonerty

An ISO 9001:2008 Registered Company
 Annacotty Business Park, Annacotty, Limerick, Ireland.
 Telephone 061 358190 Facsimile 061 358191

Hazard Identification

Labelling according to Regulation (EC) No 1272/2008
 This reagent is classed as non hazardous.