

Certificate of Performance

Phosphate 1 (Phos 1) Reagent

Product Code KR106

Lot Number 4355

Date 13.03.2020


Expiry Date 10.03.2021

Reagent	Lot Number
Phos 1	4355
Phos 2	4351
Test Method	WC SOP 21
Test Instrument	Konelab 20
Serial Number	P4919028

Standard Value	Lot #	Abs.
0		0.001
0.5	P20 03 20	0.149
1	P20 03 20	0.302
1.5	P20 03 20	0.456
2	P20 03 20	0.595

*serial dilution

Calibration Curve



	Result	Specification	Pass / Fail
Intercept	0.002	-0.005 to 0.010	Pass
Sensitivity Abs/ppm	0.299	0.260 to 0.320	Pass
Regression	1.000	0.990 to 1.001	Pass

Notes Store between 2 & 8 °C
 *Stable on Board 5 days
 Barcode ID = 106
 *Based on results obtained in our Research Facility

Alan Coonerty
 A. Coonerty

An ISO 9001:2008 Registered Company
 Annacotty Business Park, Annacotty, Limerick, Ireland.
 Telephone 061 358190 Facsimile 061 358191

Hazard Identification

Labelling according to Regulation (EC) No 1272/2008



Signal Word **Danger**

H314 Causes severe skin burns and eye damage

Precautionary statements

H314

P260; P280; P310

P303+P361+P353

P304+P340 P305+P351+P338

P321; P405; P501

see full MSDS on www.serosep.com